



Miris Calibration Control Kit™

For control of the internal calibration of the Miris Human Milk Analyzer

Miris Calibration Control Kit is a consumable to be used as a control of 'Calibration 1 - Homogenized milk' in the Miris Human Milk Analyzer (HMA).

Miris Calibration Control Kit consists of two different concentrations of a standard solution: Miris Calibration Control 1 and Miris Calibration Control 2. These are standardized solutions with known concentrations of fat, protein and carbohydrates that correspond to those found in average

human milk and the high end of the Miris Human Milk Analyzer measuring range.

The Miris Calibration Control Kit is CE-marked.

With Miris Calibration Control Kit, Miris introduces another standardization in the process of analyzing breast milk for individual nutrition for preterm babies.

MATERIAL

Environment and health Ingredients

Non-hazardous.
Aqueous mixture of fat, protein, carbohydrate, and additives.

The fat and protein are of bovine origin.

The product is retort sterilized. Non-reactive.

Storage

Store unopened vials dark, at room temperature (+20-30° C).

Do not freeze. Use opened vials directly.

Do not store opened vials to re-use the content.

Materials provided

Miris Calibration Control 1 10 pcs (8 mL)

Miris Calibration Control 2 10 pcs (8 mL)

Instructions for Use



DOC-4777 Product Sheet Miris Calibration Control kit 05-07-101 (DOC-4777)
Ver. 0

Approved By:
[\(CO-386\) CO-386 Website content and marketing material](#)

Description
Miris Website content Miris Interviews Miris Calibration Control kit product sheet

Justification
Upload of web content text and marketing material to GG

Assigned To:	Initiated By:	Priority:	Impact:
Marie Ekholm	Marie Ekholm	Low	Minor

Version History:					
Author	Effective Date	CO#	Ver.	Status	
Marie Ekholm	January 25, 2024 4:18 PM CET	CO-386	0	Published	